

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091720\
 Data File : VX018476.D
 Acq On : 17 Sep 2020 19:03
 Operator : JC/SP
 Sample : L4058-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1R

Quant Time: Sep 18 07:19:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	478740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	770538	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	735528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	359418	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	335107	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
35) Dibromofluoromethane	5.47	113	263123	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
50) Toluene-d8	8.70	98	926132	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
62) 4-Bromofluorobenzene	11.12	95	354709	46.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.14%	
Target Compounds						
16) Acetone	2.42	43	5160	1.992	ug/l #	87
19) Methyl tert-butyl Ether	3.17	73	87342	5.367	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091720\
Data File : VX018476.D
Acq On : 17 Sep 2020 19:03
Operator : JC/SP
Sample : L4058-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1R

Quant Time: Sep 18 07:19:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

