

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024307.D
 Acq On : 17 Sep 2021 14:23
 Operator : JC/MD
 Sample : M3791-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-CB-03-(1.0)

Quant Time: Sep 17 15:32:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100818	52.069	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.140%
35) Dibromofluoromethane	5.398	113	79198	50.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	8.653	98	289319	50.545	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.080%
62) 4-Bromofluorobenzene	11.086	95	83160	37.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	74.660%#
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2333	2.826	ug/l #	82
30) Chloroform	5.099	83	1352	0.485	ug/l #	77
40) Benzene	6.050	78	53034	8.584	ug/l	98
65) Chlorobenzene	10.080	112	28819	7.124	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024307.D
Acq On : 17 Sep 2021 14:23
Operator : JC/MD
Sample : M3791-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-CB-03-(1.0)

Quant Time: Sep 17 15:32:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

