

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024308.D
 Acq On : 17 Sep 2021 14:47
 Operator : JC/MD
 Sample : M3791-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-CB-02-(0.3)

Quant Time: Sep 17 15:33:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217154	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	198610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103014	53.740	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.480%	
35) Dibromofluoromethane	5.403	113	73458	50.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.880%	
50) Toluene-d8	8.653	98	292008	55.040	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 110.080%	
62) 4-Bromofluorobenzene	11.085	95	101641	49.229	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.460%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2312	2.829	ug/l #	79
30) Chloroform	5.105	83	1360	0.493	ug/l #	79
65) Chlorobenzene	10.085	112	952	0.233	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024308.D
Acq On : 17 Sep 2021 14:47
Operator : JC/MD
Sample : M3791-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-CB-02-(0.3)

Quant Time: Sep 17 15:33:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

