

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024309.D
 Acq On : 17 Sep 2021 15:10
 Operator : JC/MD
 Sample : M3791-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-01-(2.0)

Quant Time: Sep 17 15:55:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235510	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101725	52.441	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.880%	
35) Dibromofluoromethane	5.397	113	79873	50.571	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.140%	
50) Toluene-d8	8.653	98	263551	45.805	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	91.600%#	
62) 4-Bromofluorobenzene	11.085	95	92607	41.358	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	82.720%#	
Target Compounds						
16) Acetone	2.386	43	1592	1.925	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024309.D
 Acq On : 17 Sep 2021 15:10
 Operator : JC/MD
 Sample : M3791-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
Client Sampled :
 SW-RR-01-(2.0)

Quant Time: Sep 17 15:55:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

