

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024312.D
 Acq On : 17 Sep 2021 16:21
 Operator : JC/MD
 Sample : M3791-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-CB-03-(1.0)

Quant Time: Sep 17 16:43:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230263	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	237034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	106388	52.904	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.800%	
35) Dibromofluoromethane	5.397	113	82241	53.257	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.520%	
50) Toluene-d8	8.653	98	286457	50.920	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.840%	
62) 4-Bromofluorobenzene	11.085	95	115107	52.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.160%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2186	2.549	ug/l	# 87
40) Benzene	6.050	78	52690	8.678	ug/l	97
65) Chlorobenzene	10.086	112	31733	6.507	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024312.D
Acq On : 17 Sep 2021 16:21
Operator : JC/MD
Sample : M3791-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-CB-03-(1.0)

Quant Time: Sep 17 16:43:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

