

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024317.D
 Acq On : 17 Sep 2021 18:19
 Operator : JC/MD
 Sample : M3782-05 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20210914

Quant Time: Sep 20 01:59:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	218096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99884	53.274	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.540%
35) Dibromofluoromethane	5.397	113	78423	50.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.420%
50) Toluene-d8	8.653	98	284569	50.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.020%
62) 4-Bromofluorobenzene	11.085	95	98866	45.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.180%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.800	84	3282	2.138	ug/l	93
44) Trichloroethene	7.135	130	19066	11.160	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024317.D
Acq On : 17 Sep 2021 18:19
Operator : JC/MD
Sample : M3782-05 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20210914

Quant Time: Sep 20 01:59:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

