

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024318.D
 Acq On : 17 Sep 2021 18:42
 Operator : JC/MD
 Sample : M3782-06 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D6-20210914

Quant Time: Sep 20 02:00:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235034	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	190524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102016	53.024	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.040%	
35) Dibromofluoromethane	5.397	113	80111	50.824	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.640%	
50) Toluene-d8	8.653	98	244448	42.571	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 85.140%#	
62) 4-Bromofluorobenzene	11.085	95	86748	38.819	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 77.640%#	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	2692	2.027	ug/l	96
12) 1,1-Dichloroethene	2.325	96	980	0.788	ug/l	96
20) Methylene Chloride	2.800	84	2746	1.743	ug/l	96
44) Trichloroethene	7.135	130	184195	105.776	ug/l	98
64) Tetrachloroethene	9.275	164	371	0.251	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024318.D
Acq On : 17 Sep 2021 18:42
Operator : JC/MD
Sample : M3782-06 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D6-20210914

Quant Time: Sep 20 02:00:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

