

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024329.D
 Acq On : 18 Sep 2021 00:12
 Operator : JC/MD
 Sample : M3782-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D7-20210914DL

Quant Time: Sep 20 02:43:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229368	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	191717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97454	51.574	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.140%
35) Dibromofluoromethane	5.397	113	76813	49.936	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.880%
50) Toluene-d8	8.653	98	232979	41.576	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	83.160%#
62) 4-Bromofluorobenzene	11.079	95	86386	39.612	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.220%#
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	3857	2.957	ug/l	92
12) 1,1-Dichloroethene	2.325	96	1248	1.021	ug/l	82
27) cis-1,2-Dichloroethene	4.507	96	655	0.424	ug/l	69
44) Trichloroethene	7.135	130	252934	148.838	ug/l	100
64) Tetrachloroethene	9.281	164	571	0.384	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024329.D
Acq On : 18 Sep 2021 00:12
Operator : JC/MD
Sample : M3782-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D7-20210914DL

Quant Time: Sep 20 02:43:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

