

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024330.D
 Acq On : 18 Sep 2021 00:35
 Operator : JC/MD
 Sample : M3782-19 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20210915

Quant Time: Sep 20 02:43:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235712	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	193803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	100818	51.887	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.780%	
35) Dibromofluoromethane	5.397	113	79445	50.257	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.520%	
50) Toluene-d8	8.653	98	259824	45.118	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 90.240%#	
62) 4-Bromofluorobenzene	11.085	95	85536	38.167	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 76.340%#	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.331	96	378	0.301	ug/l #	45
16) Acetone	2.392	43	1271	1.534	ug/l #	82
20) Methylene Chloride	2.800	84	2312	1.453	ug/l	97
44) Trichloroethene	7.135	130	23754	13.602	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024330.D
Acq On : 18 Sep 2021 00:35
Operator : JC/MD
Sample : M3782-19 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20210915

Quant Time: Sep 20 02:43:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

