

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024331.D
 Acq On : 18 Sep 2021 00:59
 Operator : JC/MD
 Sample : M3782-20 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20210915

Quant Time: Sep 20 02:43:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	147775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	218447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110118	52.257	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.520%		
35) Dibromofluoromethane	5.397	113	85380	58.411	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 116.820%		
50) Toluene-d8	8.653	98	263314	49.449	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.900%		
62) 4-Bromofluorobenzene	11.085	95	105175	50.753	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.500%		
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	3357	2.307	ug/l	94
12) 1,1-Dichloroethene	2.325	96	2118	1.554	ug/l #	79
16) Acetone	2.398	43	1586	1.765	ug/l #	84
20) Methylene Chloride	2.794	84	3885	2.252	ug/l #	91
44) Trichloroethene	7.135	130	34021	21.068	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024331.D
Acq On : 18 Sep 2021 00:59
Operator : JC/MD
Sample : M3782-20 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20210915

Quant Time: Sep 20 02:43:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

