

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
 Data File : VX024333.D
 Acq On : 18 Sep 2021 01:46
 Operator : JC/MD
 Sample : M3782-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-8-20210914

Quant Time: Sep 20 02:44:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	195008	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	185389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	84003	51.584	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.160%	
35) Dibromofluoromethane	5.403	113	66391	50.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.540%	
50) Toluene-d8	8.653	98	237279	49.804	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.600%	
62) 4-Bromofluorobenzene	11.085	95	85541	46.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.280%	
Target Compounds						
16) Acetone	2.392	43	2068	2.978	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091721\
Data File : VX024333.D
Acq On : 18 Sep 2021 01:46
Operator : JC/MD
Sample : M3782-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-8-20210914

Quant Time: Sep 20 02:44:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

