

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047607.D
 Acq On : 17 Sep 2025 11:18
 Operator : JC/MD
 Sample : Q3095-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW305I-20250908

Quant Time: Sep 18 02:01:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	179671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	340919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	319888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	165030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	165402	57.835	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.660%
35) Dibromofluoromethane	5.373	113	117297	51.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.600%
50) Toluene-d8	8.635	98	406139	51.336	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.680%
62) 4-Bromofluorobenzene	11.061	95	166536	55.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.940%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1399	0.673	ug/l	96
12) 1,1-Dichloroethene	2.325	96	893	0.418	ug/l #	85
44) Trichloroethene	7.111	130	168949	68.867	ug/l	99
64) Tetrachloroethene	9.263	164	1622	0.778	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047607.D
Acq On : 17 Sep 2025 11:18
Operator : JC/MD
Sample : Q3095-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW305I-20250908

Quant Time: Sep 18 02:01:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

