

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047610.D
 Acq On : 17 Sep 2025 12:21
 Operator : JC/MD
 Sample : Q3095-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D1-20250908

Quant Time: Sep 18 02:02:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	175550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	325614	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	296492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	144135	51.582	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.160%		
35) Dibromofluoromethane	5.373	113	104593	47.896	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.800%		
50) Toluene-d8	8.635	98	366349	48.483	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.960%		
62) 4-Bromofluorobenzene	11.061	95	143443	50.020	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.040%		
Target Compounds						
						Qvalue
7) Trichlorofluoromethane	1.904	101	1104	0.318	ug/l #	76
12) 1,1-Dichloroethene	2.325	96	1330	0.638	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	5702	0.749	ug/l #	78
27) cis-1,2-Dichloroethene	4.483	96	37070	13.480	ug/l	89
44) Trichloroethene	7.111	130	278331	118.787	ug/l	99
64) Tetrachloroethene	9.263	164	25046	12.958	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047610.D
Acq On : 17 Sep 2025 12:21
Operator : JC/MD
Sample : Q3095-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D1-20250908

Quant Time: Sep 18 02:02:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

