

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047611.D
 Acq On : 17 Sep 2025 12:42
 Operator : JC/MD
 Sample : Q3095-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D1-20250909

Quant Time: Sep 18 02:03:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	173264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	324366	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	294888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	144060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	150415	54.539	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.080%	
35) Dibromofluoromethane	5.379	113	110705	50.890	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.780%	
50) Toluene-d8	8.635	98	386684	51.371	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.740%	
62) 4-Bromofluorobenzene	11.061	95	151703	53.104	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.200%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.331	96	12155	5.905	ug/l	92
19) Methyl tert-butyl Ether	3.111	73	6881	0.916	ug/l	98
24) 1,1-Dichloroethane	3.611	63	53206	12.159	ug/l	99
30) Chloroform	5.086	83	3770	0.853	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	7492	1.998	ug/l	98
44) Trichloroethene	7.123	130	10776	4.617	ug/l	95
64) Tetrachloroethene	9.256	164	3658	1.903	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047611.D
Acq On : 17 Sep 2025 12:42
Operator : JC/MD
Sample : Q3095-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D1-20250909

Quant Time: Sep 18 02:03:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

