

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047614.D
 Acq On : 17 Sep 2025 13:45
 Operator : JC/MD
 Sample : Q3095-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20250909

Quant Time: Sep 18 02:04:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	181930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	336757	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	299325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	150142	51.847	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.700%
35) Dibromofluoromethane	5.373	113	109089	48.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.600%
50) Toluene-d8	8.635	98	377035	48.246	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	11.061	95	144836	48.835	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.660%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.337	101	3160	1.502	ug/l	98
12) 1,1-Dichloroethene	2.331	96	620	0.287	ug/l #	82
21) trans-1,2-Dichloroethene	3.099	96	912	0.392	ug/l #	74
27) cis-1,2-Dichloroethene	4.483	96	4573	1.605	ug/l	93
44) Trichloroethene	7.110	130	605036	249.674	ug/l	98
64) Tetrachloroethene	9.256	164	3442	1.764	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047614.D
Acq On : 17 Sep 2025 13:45
Operator : JC/MD
Sample : Q3095-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20250909

Quant Time: Sep 18 02:04:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

