

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047615.D
 Acq On : 17 Sep 2025 14:06
 Operator : JC/MD
 Sample : Q3095-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D-20250909

Quant Time: Sep 18 02:04:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	176912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	325344	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	295865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	140364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	148250	52.646	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.300%
35) Dibromofluoromethane	5.373	113	107072	49.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.140%
50) Toluene-d8	8.635	98	373601	49.484	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.960%
62) 4-Bromofluorobenzene	11.061	95	144212	50.330	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.660%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.483	96	1562	0.564	ug/l	81
44) Trichloroethene	7.117	130	12485	5.333	ug/l	97
64) Tetrachloroethene	9.263	164	4397	2.280	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047615.D
Acq On : 17 Sep 2025 14:06
Operator : JC/MD
Sample : Q3095-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D-20250909

Quant Time: Sep 18 02:04:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

