

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047617.D
 Acq On : 17 Sep 2025 14:48
 Operator : JC/MD
 Sample : Q3095-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D1-20250909

Quant Time: Sep 18 02:05:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	176033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	325545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	293744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	142502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	148263	52.913	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.820%	
35) Dibromofluoromethane	5.373	113	107388	49.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.380%	
50) Toluene-d8	8.634	98	379045	50.174	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	11.067	95	145431	50.724	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.440%	
Target Compounds						Qvalue
44) Trichloroethene	7.116	130	35374	15.100	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047617.D
Acq On : 17 Sep 2025 14:48
Operator : JC/MD
Sample : Q3095-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D1-20250909

Quant Time: Sep 18 02:05:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

