

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047618.D
 Acq On : 17 Sep 2025 15:09
 Operator : JC/MD
 Sample : Q3095-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D3-20250909

Quant Time: Sep 18 02:05:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	174840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	324494	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	292220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	141405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	146146	52.514	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.020%
35) Dibromofluoromethane	5.373	113	105445	48.453	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.634	98	372448	49.461	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.067	95	145635	50.960	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.920%
Target Compounds						
						Qvalue
44) Trichloroethene	7.116	130	1307	0.560	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047618.D
Acq On : 17 Sep 2025 15:09
Operator : JC/MD
Sample : Q3095-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D3-20250909

Quant Time: Sep 18 02:05:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

