

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047624.D
 Acq On : 17 Sep 2025 17:16
 Operator : JC/MD
 Sample : Q3095-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D1-20250910

Quant Time: Sep 18 02:07:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	178815	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	331711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	303773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	145228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	146658	51.526	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.060%
35) Dibromofluoromethane	5.379	113	108715	48.869	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.740%
50) Toluene-d8	8.634	98	374888	48.701	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.067	95	144718	49.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.080%
Target Compounds						
44) Trichloroethene	7.110	130	30850	12.924	ug/l	99
64) Tetrachloroethene	9.256	164	2234	1.128	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047624.D
Acq On : 17 Sep 2025 17:16
Operator : JC/MD
Sample : Q3095-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D1-20250910

Quant Time: Sep 18 02:07:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

