

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
 Data File : VX047625.D
 Acq On : 17 Sep 2025 17:37
 Operator : JC/MD
 Sample : Q3095-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D1-20250910

Quant Time: Sep 18 02:08:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	179253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	332184	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	296569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	148841	52.165	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.340%
35) Dibromofluoromethane	5.373	113	109582	49.188	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	8.635	98	379923	49.285	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.067	95	147118	50.287	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.580%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8015	6.461	ug/l #	70
25) 2-Butanone	4.556	43	12142	7.846	ug/l #	90
29) Tetrahydrofuran	5.007	42	11097	11.920	ug/l	96
44) Trichloroethene	7.123	130	1205	0.504	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091725\
Data File : VX047625.D
Acq On : 17 Sep 2025 17:37
Operator : JC/MD
Sample : Q3095-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D1-20250910

Quant Time: Sep 18 02:08:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

