

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091818\
 Data File : VX004675.D
 Acq On : 18 Sep 2018 23:50
 Operator : JC/MD
 Sample : J4913-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW305D-20180911DL

Quant Time: Sep 19 08:36:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	337493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	320492	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	147836	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
35) Dibromofluoromethane	5.50	113	142008	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
50) Toluene-d8	8.71	98	474876	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
62) 4-Bromofluorobenzene	11.14	95	168841	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
Target Compounds						
44) Trichloroethene	7.21	130	298994	94.67	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091818\
Data File : VX004675.D
Acq On : 18 Sep 2018 23:50
Operator : JC/MD
Sample : J4913-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW305D-20180911DL

Quant Time: Sep 19 08:36:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

