

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091818\
 Data File : VX004678.D
 Acq On : 19 Sep 2018 01:10
 Operator : JC/MD
 Sample : J4938-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANKS

Quant Time: Sep 19 12:52:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	246853	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	341367	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	312744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192010	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153980	48.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.76%	
35) Dibromofluoromethane	5.50	113	142584	45.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.28%	
50) Toluene-d8	8.71	98	462555	44.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.30%	
62) 4-Bromofluorobenzene	11.14	95	161995	43.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091818\
Data File : VX004678.D
Acq On : 19 Sep 2018 01:10
Operator : JC/MD
Sample : J4938-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKS

Quant Time: Sep 19 12:52:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

