

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091818\
 Data File : VX004682.D
 Acq On : 19 Sep 2018 02:56
 Operator : JC/MD
 Sample : J4938-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9

Quant Time: Sep 19 08:53:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	343634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	313880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	198577	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	152538	47.78	ug/l	0.00
Spiked Amount			Recovery	=	95.56%	
35) Dibromofluoromethane	5.50	113	142387	45.28	ug/l	0.00
Spiked Amount			Recovery	=	90.56%	
50) Toluene-d8	8.71	98	461286	43.74	ug/l	0.00
Spiked Amount			Recovery	=	87.48%	
62) 4-Bromofluorobenzene	11.14	95	163076	43.32	ug/l	0.00
Spiked Amount			Recovery	=	86.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
19) Methyl tert-butyl Ether	3.20	73	237765	26.19	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091818\
Data File : VX004682.D
Acq On : 19 Sep 2018 02:56
Operator : JC/MD
Sample : J4938-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9

Quant Time: Sep 19 08:53:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

