

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091819\
 Data File : VX012464.D
 Acq On : 18 Sep 2019 11:31
 Operator : JC/SP
 Sample : VX0918MBL01
 Misc : 5.0µ/10mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0918MBL01

Quant Time: Sep 19 06:39:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	200203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	321121	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	271184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	108265	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	134604	46.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.82%	
35) Dibromofluoromethane	5.49	113	96918	49.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.72%	
50) Toluene-d8	8.71	98	363715	50.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	133030	44.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.08%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091819\
Data File : VX012464.D
Acq On : 18 Sep 2019 11:31
Operator : JC/SP
Sample : VX0918MBL01
Misc : 5.0g/10mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0918MBL01

Quant Time: Sep 19 06:39:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

