

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091819\
 Data File : VX012467.D
 Acq On : 18 Sep 2019 12:41
 Operator : JC/SP
 Sample : K4850-03ME
 Misc : 5.99g/5mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AOC-7(7)ME

Quant Time: Sep 18 14:19:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	282428	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	236114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	108330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	118795	46.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.32%	
35) Dibromofluoromethane	5.49	113	82676	48.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.72%	
50) Toluene-d8	8.71	98	322447	50.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.14	95	118798	45.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.44%	
Target Compounds						
40) Benzene	6.14	78	6834	1.252	ug/l	98
52) Toluene	8.78	92	5780	1.643	ug/l	100
67) Ethyl Benzene	10.25	91	1736	0.259	ug/l #	83
68) m/p-Xylenes	10.36	106	3114	1.278	ug/l	95
69) o-Xylene	10.70	106	1377	0.550	ug/l	95
80) 1,3,5-Trimethylbenzene	11.50	105	2135	0.368	ug/l	89
84) 1,2,4-Trimethylbenzene	11.81	105	3860	0.649	ug/l	94
95) Naphthalene	13.83	128	324554	43.450	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091819\
Data File : VX012467.D
Acq On : 18 Sep 2019 12:41
Operator : JC/SP
Sample : K4850-03ME
Misc : 5.99µ/5mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AOC-7(7)ME

Quant Time: Sep 18 14:19:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

