

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091819\
 Data File : VX012470.D
 Acq On : 18 Sep 2019 13:50
 Operator : JC/SP
 Sample : K4858-16DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20190910DL

Quant Time: Sep 19 06:56:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	188626	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	297724	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	255274	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	107703	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	125389	46.38	ug/l	0.00
Spiked Amount						
			Recovery	=	92.76%	
35) Dibromofluoromethane	5.49	113	88101	48.89	ug/l	0.00
Spiked Amount						
			Recovery	=	97.78%	
50) Toluene-d8	8.71	98	341340	51.18	ug/l	0.00
Spiked Amount						
			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.14	95	129974	46.93	ug/l	0.00
Spiked Amount						
			Recovery	=	93.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.36	101	1998	1.344	ug/l	94
27) cis-1,2-Dichloroethene	4.58	96	2372	1.289	ug/l	94
44) Trichloroethene	7.20	130	81502	53.847	ug/l	88
49) 1,4-Dioxane	7.76	88	477	8.340	ug/l #	69
64) Tetrachloroethene	9.33	164	5042	4.310	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091819\
Data File : VX012470.D
Acq On : 18 Sep 2019 13:50
Operator : JC/SP
Sample : K4858-16DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20190910DL

Quant Time: Sep 19 06:56:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

