

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091819\
 Data File : VX012479.D
 Acq On : 18 Sep 2019 17:21
 Operator : JC/SP
 Sample : K4858-02DL 40X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20190909DL

Quant Time: Sep 19 07:35:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	168168	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	265312	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	216704	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	82550	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	110827	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
35) Dibromofluoromethane	5.49	113	79048	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
50) Toluene-d8	8.71	98	299688	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	11.14	95	104773	42.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.90%	
Target Compounds						
44) Trichloroethene	7.21	130	46656	34.591	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091819\
Data File : VX012479.D
Acq On : 18 Sep 2019 17:21
Operator : JC/SP
Sample : K4858-02DL 40X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20190909DL

Quant Time: Sep 19 07:35:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

