

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091819\
 Data File : VX012483.D
 Acq On : 18 Sep 2019 18:54
 Operator : JC/SP
 Sample : K4858-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D1-20190910

Quant Time: Sep 19 07:45:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	189215	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	295539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244417	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126915	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
35) Dibromofluoromethane	5.49	113	88816	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
50) Toluene-d8	8.71	98	334777	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.14	95	116392	42.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.68%	
Target Compounds						
30) Chloroform	5.20	83	4882	1.370	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091819\
Data File : VX012483.D
Acq On : 18 Sep 2019 18:54
Operator : JC/SP
Sample : K4858-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D1-20190910

Quant Time: Sep 19 07:45:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

