

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091819\
 Data File : VX012485.D
 Acq On : 18 Sep 2019 19:41
 Operator : JC/SP
 Sample : K4850-03
 Misc : 5.99g/5mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AOC-7-(7)

Quant Time: Sep 19 07:50:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	175163	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	273317	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	231865	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	109721	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	113940	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	
35) Dibromofluoromethane	5.49	113	78742	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
50) Toluene-d8	8.71	98	318231	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.14	95	119501	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
Target Compounds						
40) Benzene	6.13	78	6033	1.142	ug/l	99
52) Toluene	8.78	92	5554	1.632	ug/l	99
68) m/p-Xylenes	10.36	106	2935	1.227	ug/l	89
95) Naphthalene	13.83	128	311113	41.123	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091819\
Data File : VX012485.D
Acq On : 18 Sep 2019 19:41
Operator : JC/SP
Sample : K4850-03
Misc : 5.99µ/5mL/100µL/10mL/MSVOA_X/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AOC-7-(7)

Quant Time: Sep 19 07:50:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

