

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091819\
 Data File : VX012489.D
 Acq On : 18 Sep 2019 21:14
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 19 08:01:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	118	0.00
2 T	Dichlorodifluoromethane	0.302	0.307	-1.7	105	0.00
3 P	Chloromethane	0.275	0.263	4.4	108	0.00
4 C	Vinyl Chloride	0.297	0.299	-0.7#	113	0.00
5 T	Bromomethane	0.117	0.104	11.1	111	0.00
6 T	Chloroethane	0.232	0.201	13.4	109	0.00
7 T	Trichlorofluoromethane	0.545	0.553	-1.5	112	0.00
8 T	Diethyl Ether	0.246	0.236	4.1	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.394	0.396	-0.5	115	0.00
10 T	Methyl Iodide	0.301	0.295	2.0	105	0.00
11 T	Tert butyl alcohol	0.193	0.179	7.3	113	0.00
12 CM	1,1-Dichloroethene	0.314	0.319	-1.6#	114	0.00
13 T	Acrolein	0.069	0.055	20.3#	110	0.00
14 T	Allyl chloride	0.732	0.716	2.2	109	0.00
15 T	Acrylonitrile	0.311	0.309	0.6	113	0.00
16 T	Acetone	0.363	0.284	21.8#	96	0.00
17 T	Carbon Disulfide	0.389	0.402	-3.3	106	0.00
18 T	Methyl Acetate	0.747	0.731	2.1	112	0.00
19 T	Methyl tert-butyl Ether	1.714	1.713	0.1	115	0.00
20 T	Methylene Chloride	0.434	0.427	1.6	116	0.00
21 T	trans-1,2-Dichloroethene	0.326	0.330	-1.2	113	0.00
22 T	Diisopropyl ether	1.714	1.692	1.3	113	0.00
23 T	Vinyl Acetate	1.410	1.435	-1.8	112	0.00
24 P	1,1-Dichloroethane	0.820	0.834	-1.7	114	0.00
25 T	2-Butanone	0.497	0.458	7.8	107	0.00
26 T	2,2-Dichloropropane	0.806	0.668	17.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.488	0.491	-0.6	116	0.00
28 T	Bromochloromethane	0.423	0.363	14.2	106	0.00
29 T	Tetrahydrofuran	0.275	0.279	-1.5	112	0.00
30 C	Chloroform	0.942	0.915	2.9#	114	0.00
31 T	Cyclohexane	0.495	0.509	-2.8	114	0.00
32 T	1,1,1-Trichloroethane	0.787	0.784	0.4	113	0.00
33 S	1,2-Dichloroethane-d4	0.717	0.688	4.0	124	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	118	0.00
35 S	Dibromofluoromethane	0.303	0.304	-0.3	127	0.00
36 T	1,1-Dichloropropene	0.304	0.301	1.0	111	0.00
37 T	Ethyl Acetate	0.466	0.488	-4.7	111	0.00
38 T	Carbon Tetrachloride	0.378	0.381	-0.8	113	0.00
39 T	Methylcyclohexane	0.295	0.301	-2.0	108	0.00
40 TM	Benzene	0.966	0.992	-2.7	116	0.00
41 T	Methacrylonitrile	0.300	0.285	5.0	109	0.00
42 TM	1,2-Dichloroethane	0.448	0.435	2.9	112	0.00
43 T	Isopropyl Acetate	0.837	0.847	-1.2	113	0.00
44 TM	Trichloroethene	0.254	0.261	-2.8	113	0.00
45 C	1,2-Dichloropropane	0.297	0.298	-0.3#	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091819\
 Data File : VX012489.D
 Acq On : 18 Sep 2019 21:14
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 19 08:01:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.192	0.195	-1.6	115	0.00
47 T	Bromodichloromethane	0.428	0.433	-1.2	112	0.00
48 T	Methyl methacrylate	0.395	0.390	1.3	109	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	115	0.00
50 S	Toluene-d8	1.120	1.123	-0.3	128	0.00
51 T	4-Methyl-2-Pentanone	0.549	0.549	0.0	111	0.00
52 CM	Toluene	0.623	0.644	-3.4#	115	0.00
53 T	t-1,3-Dichloropropene	0.445	0.456	-2.5	111	0.00
54 T	cis-1,3-Dichloropropene	0.460	0.477	-3.7	112	0.00
55 T	1,1,2-Trichloroethane	0.321	0.322	-0.3	116	0.00
56 T	Ethyl methacrylate	0.502	0.521	-3.8	115	0.00
57 T	1,3-Dichloropropane	0.512	0.520	-1.6	114	0.00
58 T	2-Chloroethyl Vinyl ether	0.260	0.280	-7.7	124	0.00
59 T	2-Hexanone	0.428	0.422	1.4	109	0.00
60 T	Dibromochloromethane	0.327	0.351	-7.3	115	0.00
61 T	1,2-Dibromoethane	0.293	0.317	-8.2	117	0.00
62 S	4-Bromofluorobenzene	0.465	0.453	2.6	125	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	118	0.00
64 T	Tetrachloroethene	0.229	0.252	-10.0	122	0.00
65 PM	Chlorobenzene	0.825	0.844	-2.3	117	0.00
66 T	1,1,1,2-Tetrachloroethane	0.342	0.366	-7.0	117	0.00
67 C	Ethyl Benzene	1.418	1.467	-3.5#	114	0.00
68 T	m/p-Xylenes	0.516	0.539	-4.5	115	0.00
69 T	o-Xylene	0.531	0.562	-5.8	116	0.00
70 T	Styrene	0.939	1.002	-6.7	116	0.00
71 P	Bromoform	0.256	0.282	-10.2	116	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
73 T	Isopropylbenzene	3.137	3.165	-0.9	113	0.00
74 T	N-amyl acetate	1.689	1.743	-3.2	113	0.00
75 P	1,1,2,2-Tetrachloroethane	1.231	1.232	-0.1	114	0.00
76 T	1,2,3-Trichloropropane	1.149	1.162	-1.1	111	0.00
77 T	Bromobenzene	0.765	0.802	-4.8	118	0.00
78 T	n-propylbenzene	3.497	3.555	-1.7	113	0.00
79 T	2-Chlorotoluene	2.275	2.223	2.3	113	0.00
80 T	1,3,5-Trimethylbenzene	2.676	2.727	-1.9	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.387	0.366	5.4	103	0.00
82 T	4-Chlorotoluene	2.611	2.605	0.2	113	0.00
83 T	tert-Butylbenzene	2.867	2.844	0.8	113	0.00
84 T	1,2,4-Trimethylbenzene	2.747	2.830	-3.0	115	0.00
85 T	sec-Butylbenzene	3.159	3.193	-1.1	112	0.00
86 T	p-Isopropyltoluene	2.901	2.952	-1.8	112	0.00
87 T	1,3-Dichlorobenzene	1.439	1.485	-3.2	117	0.00
88 T	1,4-Dichlorobenzene	1.478	1.516	-2.6	118	0.00
89 T	n-Butylbenzene	2.588	2.552	1.4	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091819\
Data File : VX012489.D
Acq On : 18 Sep 2019 21:14
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 19 08:01:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.532	0.541	-1.7	110	0.00
91 T	1,2-Dichlorobenzene	1.504	1.566	-4.1	117	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.319	0.327	-2.5	110	0.00
93 T	1,2,4-Trichlorobenzene	0.969	1.053	-8.7	117	0.00
94 T	Hexachlorobutadiene	0.457	0.434	5.0	108	0.00
95 T	Naphthalene	3.448	3.891	-12.8	119	0.00
96 T	1,2,3-Trichlorobenzene	0.987	1.095	-10.9	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6