

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091820\
 Data File : VX018484.D
 Acq On : 18 Sep 2020 11:48
 Operator : JC/SP
 Sample : VX0918MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0918MBL01

Quant Time: Sep 21 01:23:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	451132	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	727754	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	676104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	328133	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	316814	49.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.28%	
35) Dibromofluoromethane	5.47	113	248864	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
50) Toluene-d8	8.70	98	864279	47.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.56%	
62) 4-Bromofluorobenzene	11.12	95	315605	43.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.80%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091820\
Data File : VX018484.D
Acq On : 18 Sep 2020 11:48
Operator : JC/SP
Sample : VX0918MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0918MBL01

Quant Time: Sep 21 01:23:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

