

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091820\
 Data File : VX018498.D
 Acq On : 18 Sep 2020 17:26
 Operator : JC/SP
 Sample : L4058-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Sep 21 07:22:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	552510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	901996	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	844672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	431987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	373625	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
35) Dibromofluoromethane	5.47	113	296608	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
50) Toluene-d8	8.70	98	1083630	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
62) 4-Bromofluorobenzene	11.12	95	410140	45.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.02%	
Target Compounds						
16) Acetone	2.42	43	15760	5.272	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091820\
Data File : VX018498.D
Acq On : 18 Sep 2020 17:26
Operator : JC/SP
Sample : L4058-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Sep 21 07:22:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

