

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091820\
 Data File : VX018503.D
 Acq On : 18 Sep 2020 19:19
 Operator : JC/SP
 Sample : L4058-06DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7DL

Quant Time: Sep 21 03:49:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	515406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	824742	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	779055	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	377188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	347627	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
35) Dibromofluoromethane	5.46	113	276255	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	8.70	98	1000323	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	11.12	95	370650	44.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.96%	
Target Compounds						
11) Tert butyl alcohol	3.01	59	470792	371.100	ug/l	99
19) Methyl tert-butyl Ether	3.17	73	1039569	59.330	ug/l	100
22) Diisopropyl ether	3.83	45	17471	0.961	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091820\
Data File : VX018503.D
Acq On : 18 Sep 2020 19:19
Operator : JC/SP
Sample : L4058-06DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7DL

Quant Time: Sep 21 03:49:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

