


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091820\
Data File : VX018479.D
Acq On : 18 Sep 2020 09:21
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Title : SW846 8260
Last Update : Thu Sep 17 12:17:36 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	21994	PASS
75	95	30	60	50.7	64352	PASS
95	95	100	100	100.0	126890	PASS
96	95	5	9	6.3	8009	PASS
173	174	0.00	2	1.9	1979	PASS
174	95	50	100	80.5	102152	PASS
175	174	5	9	7.5	7709	PASS
176	174	95	101	98.2	100357	PASS
177	176	5	9	6.9	6930	PASS