

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037767.D
 Acq On : 18 Sep 2023 10:34
 Operator : JC/MD
 Sample : VX0918MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0918MBL01

Quant Time: Sep 18 17:54:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	228585	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	372428	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	325675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	133879	44.578	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.160%
35) Dibromofluoromethane	5.385	113	116044	46.568	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.140%
50) Toluene-d8	8.647	98	439998	47.521	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.040%
62) 4-Bromofluorobenzene	11.079	95	146798	43.630	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037767.D
Acq On : 18 Sep 2023 10:34
Operator : JC/MD
Sample : VX0918MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0918MBL01

Quant Time: Sep 18 17:54:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

