

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037773.D
 Acq On : 18 Sep 2023 13:44
 Operator : JC/MD
 Sample : 04437-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-220-222

Quant Time: Sep 18 17:56:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	297909	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	491672	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	456495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	217648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	178943	45.718	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.440%
35) Dibromofluoromethane	5.391	113	153443	46.642	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	8.653	98	585678	47.914	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.820%
62) 4-Bromofluorobenzene	11.079	95	216554	48.753	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.500%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5383	2.859	ug/l	# 78
19) Methyl tert-butyl Ether	3.111	73	2876	0.269	ug/l	100
22) Diisopropyl ether	3.757	45	6574	0.652	ug/l	# 67
24) 1,1-Dichloroethane	3.611	63	2135	0.353	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037773.D
Acq On : 18 Sep 2023 13:44
Operator : JC/MD
Sample : 04437-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-220-222

Quant Time: Sep 18 17:56:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

