

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037776.D
 Acq On : 18 Sep 2023 14:54
 Operator : JC/MD
 Sample : 04437-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-260-262

Quant Time: Sep 18 17:57:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	281225	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	468308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	440234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	221853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	172231	46.614	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.220%
35) Dibromofluoromethane	5.391	113	148768	47.477	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.960%
50) Toluene-d8	8.653	98	560955	48.180	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.079	95	211853	50.074	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.140%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3532	1.987	ug/l	95
17) Carbon Disulfide	2.514	76	2782	0.404	ug/l #	93
22) Diisopropyl ether	3.770	45	4731	0.497	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037776.D
Acq On : 18 Sep 2023 14:54
Operator : JC/MD
Sample : 04437-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-260-262

Quant Time: Sep 18 17:57:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

