

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037779.D
 Acq On : 18 Sep 2023 16:04
 Operator : JC/MD
 Sample : 04437-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-300-302

Quant Time: Sep 18 17:57:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	316964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	528428	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	495872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	245964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	192348	46.188	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.380%		
35) Dibromofluoromethane	5.391	113	165169	46.714	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.420%		
50) Toluene-d8	8.647	98	626921	47.720	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.440%		
62) 4-Bromofluorobenzene	11.079	95	233306	48.871	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.740%		
Target Compounds						
16) Acetone	2.379	43	3344	1.669	ug/l #	86
17) Carbon Disulfide	2.514	76	3406	0.439	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037779.D
Acq On : 18 Sep 2023 16:04
Operator : JC/MD
Sample : 04437-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-300-302

Quant Time: Sep 18 17:57:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

