

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037786.D
 Acq On : 18 Sep 2023 19:10
 Operator : JC/MD
 Sample : 04389-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20230907

Quant Time: Sep 19 07:25:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	293210	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	495768	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	473624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	243905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	184961	48.013	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.020%
35) Dibromofluoromethane	5.391	113	156826	47.276	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.560%
50) Toluene-d8	8.647	98	603057	48.928	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	227523	50.799	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.600%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	384287	129.834	ug/l	96
12) 1,1-Dichloroethene	2.325	96	2976	1.040	ug/l	# 71
27) cis-1,2-Dichloroethene	4.507	96	2926	0.751	ug/l	77
44) Trichloroethene	7.129	130	81868	20.920	ug/l	97
64) Tetrachloroethene	9.275	164	43207	13.377	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037786.D
Acq On : 18 Sep 2023 19:10
Operator : JC/MD
Sample : 04389-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20230907

Quant Time: Sep 19 07:25:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

