

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037788.D
 Acq On : 18 Sep 2023 19:56
 Operator : JC/MD
 Sample : 04408-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14-0912

Quant Time: Sep 19 07:26:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	243926	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	404235	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	392590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	203218	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147945	46.163	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.320%
35) Dibromofluoromethane	5.391	113	124921	46.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.653	98	493701	49.125	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	186253	51.001	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037788.D
Acq On : 18 Sep 2023 19:56
Operator : JC/MD
Sample : 04408-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14-0912

Quant Time: Sep 19 07:26:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

