

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
 Data File : VX037789.D
 Acq On : 18 Sep 2023 20:19
 Operator : JC/MD
 Sample : 04459-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137BASIN-A-20230914

Quant Time: Sep 19 07:26:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	269357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	450934	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	415523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	203502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166748	47.118	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.240%
35) Dibromofluoromethane	5.391	113	141532	46.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.820%
50) Toluene-d8	8.653	98	535537	47.770	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.540%
62) 4-Bromofluorobenzene	11.079	95	194518	47.748	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.500%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.337	101	3707	1.363	ug/l	95
16) Acetone	2.380	43	2198	1.291	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091823\
Data File : VX037789.D
Acq On : 18 Sep 2023 20:19
Operator : JC/MD
Sample : 04459-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137BASIN-A-20230914

Quant Time: Sep 19 07:26:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

