

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
 Data File : VX047636.D
 Acq On : 18 Sep 2025 13:37
 Operator : JC/MD
 Sample : Q3095-11DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20250909DL

Quant Time: Sep 19 01:30:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	174332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	317792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	290497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	143589	51.745	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.500%
35) Dibromofluoromethane	5.373	113	106906	50.160	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.320%
50) Toluene-d8	8.635	98	373354	50.626	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.260%
62) 4-Bromofluorobenzene	11.061	95	144794	51.734	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.460%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.343	101	733	0.364	ug/l #	85
44) Trichloroethene	7.111	130	122156	53.417	ug/l	100
64) Tetrachloroethene	9.256	164	838	0.443	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
Data File : VX047636.D
Acq On : 18 Sep 2025 13:37
Operator : JC/MD
Sample : Q3095-11DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20250909DL

Quant Time: Sep 19 01:30:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

