

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
 Data File : VX047638.D
 Acq On : 18 Sep 2025 14:19
 Operator : JC/MD
 Sample : Q3095-13DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17811-20250909DL

Quant Time: Sep 19 01:31:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	192027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	349705	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	314014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	151081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	160540	52.523	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.040%			
35) Dibromofluoromethane	5.373	113	118528	50.538	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.080%			
50) Toluene-d8	8.634	98	417494	51.446	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.900%			
62) 4-Bromofluorobenzene	11.061	95	161415	52.409	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.820%			
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.489	96	7246	2.409	ug/l	94
44) Trichloroethene	7.110	130	92416	36.724	ug/l	100
45) 1,2-Dichloropropane	7.415	63	5075	1.904	ug/l	93
64) Tetrachloroethene	9.262	164	6200	3.029	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
Data File : VX047638.D
Acq On : 18 Sep 2025 14:19
Operator : JC/MD
Sample : Q3095-13DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17811-20250909DL

Quant Time: Sep 19 01:31:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

