

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
 Data File : VX047641.D
 Acq On : 18 Sep 2025 15:22
 Operator : JC/MD
 Sample : Q3095-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D-20250909

Quant Time: Sep 19 01:32:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	185198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	333961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	301489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	144225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	162054	54.973	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.940%	
35) Dibromofluoromethane	5.379	113	122491	54.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.380%	
50) Toluene-d8	8.634	98	423743	54.677	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.360%	
62) 4-Bromofluorobenzene	11.061	95	163779	55.684	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.360%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.477	96	1620	0.558	ug/l	90
44) Trichloroethene	7.116	130	11063	4.603	ug/l	94
64) Tetrachloroethene	9.262	164	4258	2.166	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
Data File : VX047641.D
Acq On : 18 Sep 2025 15:22
Operator : JC/MD
Sample : Q3095-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D-20250909

Quant Time: Sep 19 01:32:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

