

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
 Data File : VX047656.D
 Acq On : 18 Sep 2025 20:34
 Operator : JC/MD
 Sample : Q3095-31
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D-20250911

Quant Time: Sep 19 01:37:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	161049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	300961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	275687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	133701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	161876	63.147	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 126.300%#	
35) Dibromofluoromethane	5.379	113	119338	59.125	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 118.240%	
50) Toluene-d8	8.635	98	417521	59.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 119.560%#	
62) 4-Bromofluorobenzene	11.061	95	161244	60.833	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 121.660%	
Target Compounds						
44) Trichloroethene	7.117	130	10643	4.914	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
Data File : VX047656.D
Acq On : 18 Sep 2025 20:34
Operator : JC/MD
Sample : Q3095-31
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D-20250911

Quant Time: Sep 19 01:37:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

