

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
 Data File : VX047657.D
 Acq On : 18 Sep 2025 20:55
 Operator : JC/MD
 Sample : Q3095-32
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20250911

Quant Time: Sep 19 01:37:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	166478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	307321	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	278614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	142746	53.868	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.740%	
35) Dibromofluoromethane	5.379	113	105031	50.960	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.920%	
50) Toluene-d8	8.635	98	368111	51.616	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.240%	
62) 4-Bromofluorobenzene	11.061	95	141338	52.220	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.440%	
Target Compounds						
					Qvalue	
44) Trichloroethene	7.117	130	8121	3.672	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091825\
Data File : VX047657.D
Acq On : 18 Sep 2025 20:55
Operator : JC/MD
Sample : Q3095-32
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20250911

Quant Time: Sep 19 01:37:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

