

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004719.D
 Acq On : 19 Sep 2018 21:13
 Operator : JC/MD
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :

Quant Time: Sep 20 03:52:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	214497	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	302416	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	177116	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	119725	43.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.58%	
35) Dibromofluoromethane	5.51	113	109897	39.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.42%	
50) Toluene-d8	8.72	98	411748	44.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.74%	
62) 4-Bromofluorobenzene	11.14	95	153280	46.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.54%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.19	85	104499	51.98	ug/l	98
3) Chloromethane	1.32	50	140511	50.79	ug/l	96
4) Vinyl Chloride	1.40	62	155307	54.70	ug/l	98
5) Bromomethane	1.64	94	66441	56.03	ug/l	98
6) Chloroethane	1.71	64	95111	51.80	ug/l	95
7) Trichlorofluoromethane	1.92	101	201318	53.48	ug/l	97
8) Diethyl Ether	2.19	74	91753	54.37	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	2.38	101	125142	52.41	ug/l	97
10) Methyl Iodide	2.51	142	140114	49.16	ug/l	97
11) Tert butyl alcohol	3.08	59	188092	285.01	ug/l	99
12) 1,1-Dichloroethene	2.37	96	124635	55.61	ug/l	90
13) Acrolein	2.29	56	51744	198.63	ug/l	95
14) Allyl chloride	2.72	41	241606	52.44	ug/l	95
15) Acrylonitrile	3.15	53	440310	277.06	ug/l	99
16) Acetone	2.45	43	330023	244.48	ug/l	94
17) Carbon Disulfide	2.56	76	313179	48.34	ug/l	99
18) Methyl Acetate	2.78	43	228143	60.24	ug/l	98
19) Methyl tert-butyl Ether	3.20	73	441382	56.13	ug/l	96
20) Methylene Chloride	2.85	84	144426	52.16	ug/l	92
21) trans-1,2-Dichloroethene	3.16	96	130825	52.82	ug/l	97
22) Diisopropyl ether	3.87	45	456415	55.00	ug/l	96
23) Vinyl Acetate	3.82	43	1772276	263.90	ug/l	99
24) 1,1-Dichloroethane	3.70	63	245307	52.36	ug/l	99
25) 2-Butanone	4.70	43	528991	246.23	ug/l	94
26) 2,2-Dichloropropane	4.58	77	150769	46.85	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	147829	53.47	ug/l	100
28) Bromochloromethane	5.01	49	97000	45.26	ug/l	98
29) Tetrahydrofuran	5.15	42	346989	267.82	ug/l	100
30) Chloroform	5.21	83	245932	53.76	ug/l	94
31) Cyclohexane	5.57	56	204371	51.08	ug/l	96
32) 1,1,1-Trichloroethane	5.50	97	204235	55.44	ug/l	98
36) 1,1-Dichloropropene	5.80	75	168008	45.76	ug/l	98
37) Ethyl Acetate	4.86	43	203302	48.69	ug/l	99
38) Carbon Tetrachloride	5.78	117	170075	47.43	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004719.D
 Acq On : 19 Sep 2018 21:13
 Operator : JC/MD
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 20 03:52:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	201538	52.60	ug/l	98
40) Benzene	6.14	78	523835	48.09	ug/l	100
41) Methacrylonitrile	5.06	41	111708	48.55	ug/l	97
42) 1,2-Dichloroethane	6.20	62	174230	48.22	ug/l	99
43) Isopropyl Acetate	6.46	43	293368	52.90	ug/l	99
44) Trichloroethene	7.21	130	156901	55.44	ug/l	98
45) 1,2-Dichloropropane	7.52	63	139438	53.20	ug/l	100
46) Dibromomethane	7.66	93	89128	53.82	ug/l	98
47) Bromodichloromethane	7.90	83	169803	54.22	ug/l	99
48) Methyl methacrylate	7.77	41	148406	54.07	ug/l	97
49) 1,4-Dioxane	7.76	88	75103	1156.07	ug/l	98
51) 4-Methyl-2-Pentanone	8.65	43	973936	248.84	ug/l	99
52) Toluene	8.79	92	332880	51.87	ug/l	97
53) t-1,3-Dichloropropene	9.04	75	186397	50.95	ug/l	99
54) cis-1,3-Dichloropropene	8.44	75	206234	55.30	ug/l	95
55) 1,1,2-Trichloroethane	9.22	97	138776	50.22	ug/l	97
56) Ethyl methacrylate	9.18	69	209245	53.52	ug/l	96
57) 1,3-Dichloropropane	9.37	76	226085	50.35	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.32	63	482476	242.45	ug/l	98
59) 2-Hexanone	9.50	43	741175	246.88	ug/l	100
60) Dibromochloromethane	9.59	129	142389	54.73	ug/l	100
61) 1,2-Dibromoethane	9.68	107	141992	53.79	ug/l	100
64) Tetrachloroethene	9.34	164	191452	53.54	ug/l	97
65) Chlorobenzene	10.14	112	384172	54.00	ug/l	95
66) 1,1,1,2-Tetrachloroethane	10.23	131	138088	56.28	ug/l	99
67) Ethyl Benzene	10.26	91	639463	55.27	ug/l	100
68) m/p-Xylenes	10.36	106	505114	113.35	ug/l	99
69) o-Xylene	10.70	106	246200	57.47	ug/l	99
70) Styrene	10.71	104	415766	58.58	ug/l	99
71) Bromoform	10.86	173	114872	57.02	ug/l	99
73) Isopropylbenzene	11.02	105	661771	56.73	ug/l	100
74) N-amyl acetate	10.90	43	264912	53.96	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.27	83	188955	50.22	ug/l	100
76) 1,2,3-Trichloropropane	11.30	75	243707	68.39	ug/l	86
77) Bromobenzene	11.26	156	180445	55.97	ug/l	97
78) n-propylbenzene	11.36	91	751693	56.05	ug/l	99
79) 2-Chlorotoluene	11.42	91	442149	54.36	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	557906	56.89	ug/l	98
81) trans-1,4-Dichloro-2-buten	11.08	75	60105	53.76	ug/l	94
82) 4-Chlorotoluene	11.51	91	518748	54.64	ug/l	99
83) tert-Butylbenzene	11.77	119	527988	56.56	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	571917	57.12	ug/l	99
85) sec-Butylbenzene	11.95	105	678769	57.06	ug/l	100
86) p-Isopropyltoluene	12.07	119	608593	57.93	ug/l	100
87) 1,3-Dichlorobenzene	12.03	146	329191	54.72	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	331749	53.37	ug/l	100
89) n-Butylbenzene	12.39	91	521800	54.51	ug/l	98
90) Hexachloroethane	12.60	117	95099	56.55	ug/l	93
91) 1,2-Dichlorobenzene	12.40	146	335875	55.16	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.00	75	45345	56.44	ug/l	95

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004719.D
 Acq On : 19 Sep 2018 21:13
 Operator : JC/MD
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 20 03:52:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	250594	58.13	ug/l	98
94) Hexachlorobutadiene	13.79	225	123598	54.88	ug/l	98
95) Naphthalene	13.83	128	731376	60.49	ug/l	100
96) 1,2,3-Trichlorobenzene	14.02	180	255542	57.70	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004719.D
 Acq On : 19 Sep 2018 21:13
 Operator : JC/MD
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sampled :

Quant Time: Sep 20 03:52:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration









#1
 Pentafluorobenzene
 Concen: 50.00 ug/l
 RT: 5.67 min Scan# 740
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
168	100		
99	43.6	39.9	59.9



#2
 Dichlorodifluoromethane
 Concen: 51.98 ug/l
 RT: 1.19 min Scan# 6
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
85	100		
87	34.8	17.0	50.9









#3
 Chloromethane
 Concen: 50.79 ug/l
 RT: 1.32 min Scan# 28
 Delta R.T. 0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
50	140511		
52	30.4	26.2	39.2



#4
 Vinyl Chloride
 Concen: 54.70 ug/l
 RT: 1.40 min Scan# 41
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
62	155307		
64	33.5	25.8	38.8









#5
 Bromomethane
 Concen: 56.03 ug/l
 RT: 1.64 min Scan# 79
 Delta R.T. 0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
94	66441		
96	94.7	74.5	111.7



#6
 Chloroethane
 Concen: 51.80 ug/l
 RT: 1.71 min Scan# 92
 Delta R.T. 0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
64	95111		
66	33.4	24.6	36.8









#7
 Trichlorofluoromethane
 Concen: 53.48 ug/l
 RT: 1.92 min Scan# 126
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
101	201318		
103	63.7	48.9	73.3



#8
 Diethyl Ether
 Concen: 54.37 ug/l
 RT: 2.19 min Scan# 170
 Delta R.T. 0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
74	91753		
45	98.2	48.3	144.8









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 52.41 ug/l
 RT: 2.38 min Scan# 201
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
101	125142		
85	42.6	34.2	51.4
151	85.3	65.0	97.6



#10
 Methyl Iodide
 Concen: 49.16 ug/l
 RT: 2.51 min Scan# 222
 Delta R.T. 0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
142	140114		
127	39.8	33.8	50.6
141	14.0	11.4	17.0









#11
 Tert butyl alcohol
 Concen: 285.01 ug/l
 RT: 3.08 min Scan# 316
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
59	100		
57	9.9	8.2	12.4



#12
 1,1-Dichloroethene
 Concen: 55.61 ug/l
 RT: 2.37 min Scan# 199
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
96	100		
61	146.2	128.8	193.2
98	60.0	52.2	78.2









#13
 Acrolein
 Concen: 198.63 ug/l
 RT: 2.29 min Scan# 187
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
56	51744		
55	72.3	54.7	82.1



#14
 Allyl chloride
 Concen: 52.44 ug/l
 RT: 2.72 min Scan# 257
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
41	241606		
39	55.9	48.9	73.3
76	32.0	26.7	40.1









#15
 Acrylonitrile
 Concen: 277.06 ug/l
 RT: 3.15 min Scan# 327
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
53	440310		
52	81.3	65.2	97.8
51	33.7	28.2	42.2



#16
 Acetone
 Concen: 244.48 ug/l
 RT: 2.45 min Scan# 212
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
43	330023		
58	35.9	26.2	39.4









#17
 Carbon Disulfide
 Concen: 48.34 ug/l
 RT: 2.56 min Scan# 231
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
76	313179		
78	100	7.0	10.6



#18
 Methyl Acetate
 Concen: 60.24 ug/l
 RT: 2.78 min Scan# 266
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
43	228143		
74	100	19.4	29.2









#19
 Methyl tert-butyl Ether
 Concen: 56.13 ug/l
 RT: 3.20 min Scan# 336
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
73	441382		
57	23.1	17.1	25.7



#20
 Methylene Chloride
 Concen: 52.16 ug/l
 RT: 2.85 min Scan# 279
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
84	144426		
49	114.4	101.2	151.8
51	35.9	29.5	44.3
86	59.8	52.3	78.5









#21
 trans-1,2-Dichloroethene
 Concen: 52.82 ug/l
 RT: 3.16 min Scan# 329
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
96	130825		
96	100		
61	136.1	113.8	170.6
98	64.5	51.8	77.8



#22
 Diisopropyl ether
 Concen: 55.00 ug/l
 RT: 3.87 min Scan# 446
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
45	456415		
45	100		
43	50.8	42.8	64.2
87	31.5	22.6	34.0
59	11.7	9.6	14.4









#23
 Vinyl Acetate
 Concen: 263.90 ug/l
 RT: 3.82 min Scan# 437
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Instrument :
 MSVOA_X
 ClientSampled :

Tgt Ion: 43 Resp: 1772276




Ion	Ratio	Lower	Upper
43	100		
86	10.5	8.9	13.3




#24
 1,1-Dichloroethane
 Concen: 52.36 ug/l
 RT: 3.70 min Scan# 417
 Delta R.T. 0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion: 63 Resp: 245307

Ion	Ratio	Lower	Upper
63	100		
98	7.4	3.5	10.4
100	4.7	2.4	7.1









#25
 2-Butanone
 Concen: 246.23 ug/l
 RT: 4.70 min Scan# 582
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
43	100		
72	24.2	22.0	33.0



#26
 2,2-Dichloropropane
 Concen: 46.85 ug/l
 RT: 4.58 min Scan# 562
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
77	100		
97	25.0	11.7	35.0









#27
 cis-1,2-Dichloroethene
 Concen: 53.47 ug/l
 RT: 4.60 min Scan# 565
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
96	147829		
61	141.2	0.0	282.6
98	65.2	0.0	130.2



#28
 Bromochloromethane
 Concen: 45.26 ug/l
 RT: 5.01 min Scan# 633
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
49	97000		
129	2.0	0.0	4.2
130	86.1	67.5	101.3









#29
 Tetrahydrofuran
 Concen: 267.82 ug/l
 RT: 5.15 min Scan# 656
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
42	346989		
72	46.5	37.4	56.0
71	42.8	34.5	51.7



#30
 Chloroform
 Concen: 53.76 ug/l
 RT: 5.21 min Scan# 666
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
83	245932		
85	60.9	52.6	79.0









#31
 Cyclohexane
 Concen: 51.08 ug/l
 RT: 5.57 min Scan# 725
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
56	100		
69	29.9	25.4	38.2
84	85.4	71.4	107.2



#32
 1,1,1-Trichloroethane
 Concen: 55.44 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
97	100		
99	64.3	52.0	78.0
61	40.6	34.9	52.3









#33
 1,2-Dichloroethane-d4
 Concen: 43.29 ug/l
 RT: 6.07 min Scan# 806
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
65	119725		
67	55.7	0.0	106.8



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
114	302416		
63	18.7	0.0	39.0
88	15.9	0.0	30.4









#35
 Dibromofluoromethane
 Concen: 39.71 ug/l
 RT: 5.51 min Scan# 714
 Delta R.T. 0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
113	109897		
111	104.2	82.4	123.6
192	24.8	19.4	29.2



#36
 1,1-Dichloropropene
 Concen: 45.76 ug/l
 RT: 5.80 min Scan# 762
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
75	168008		
110	37.8	18.0	54.0
77	31.5	24.6	36.8









#37
 Ethyl Acetate
 Concen: 48.69 ug/l
 RT: 4.86 min Scan# 608
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
43	100		
61	14.5	11.5	17.3
70	10.8	9.0	13.6



#38
 Carbon Tetrachloride
 Concen: 47.43 ug/l
 RT: 5.78 min Scan# 759
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
117	100		
119	95.0	78.8	118.2
121	30.8	24.9	37.3









#39
 Methylcyclohexane
 Concen: 52.60 ug/l
 RT: 7.46 min Scan# 1035
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
83	100		
55	75.8	61.0	91.4
98	46.3	39.6	59.4



#40
 Benzene
 Concen: 48.09 ug/l
 RT: 6.14 min Scan# 818
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
78	100		
77	23.6	19.0	28.4









#41
 Methacrylonitrile
 Concen: 48.55 ug/l
 RT: 5.06 min Scan# 640
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Instrument :
 MSVOA_X
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
41	111708		
39	52.1	42.4	63.6
67	77.5	59.2	88.8
52	29.6	23.0	34.4




Abundance Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0
 Ion 67.00 (66.70 to 67.70): VX0
 Ion 52.00 (51.70 to 52.70): VX0






#42
 1,2-Dichloroethane
 Concen: 48.22 ug/l
 RT: 6.20 min Scan# 827
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
62	174230		
98	10.0	0.0	20.6



Abundance Ion 61.90 (61.60 to 62.60): VX0
 Ion 97.90 (97.60 to 98.60): VX0









#43
 Isopropyl Acetate
 Concen: 52.90 ug/l
 RT: 6.46 min Scan# 871
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
43	100		
61	21.0	17.0	25.4
87	13.8	11.4	17.2



#44
 Trichloroethene
 Concen: 55.44 ug/l
 RT: 7.21 min Scan# 994
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
130	100		
95	92.9	0.0	181.6









#45
 1,2-Dichloropropane
 Concen: 53.20 ug/l
 RT: 7.52 min Scan# 1044
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
63	139438		
65	30.4	24.3	36.5



#46
 Dibromomethane
 Concen: 53.82 ug/l
 RT: 7.66 min Scan# 1068
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
93	89128		
95	83.8	65.5	98.3
174	120.8	94.2	141.4








#47
 Bromodichloromethane
 Concen: 54.22 ug/l
 RT: 7.90 min Scan# 1107
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13





Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
83	169803		
85	64.7	50.9	76.3
127	9.5	7.3	10.9



#48
 Methyl methacrylate
 Concen: 54.07 ug/l
 RT: 7.77 min Scan# 1086
 Delta R.T. -0.01 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
41	148406		
69	89.1	70.2	105.4
39	49.6	42.7	64.1









#49
 1,4-Dioxane
 Concen: 1156.07 ug/l
 RT: 7.76 min Scan# 1083
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
88	75103		
43	28.1	24.7	37.1
58	68.7	55.3	82.9



#50
 Toluene-d8
 Concen: 44.37 ug/l
 RT: 8.72 min Scan# 1241
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
98	411748		
100	64.8	52.9	79.3









#51
 4-Methyl-2-Pentanone
 Concen: 248.84 ug/l
 RT: 8.65 min Scan# 1230
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
43	100		
58	40.9	33.1	49.7



#52
 Toluene
 Concen: 51.87 ug/l
 RT: 8.79 min Scan# 1252
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
92	100		
91	174.0	136.4	204.6









#53
 t-1,3-Dichloropropene
 Concen: 50.95 ug/l
 RT: 9.04 min Scan# 1294
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
75	186397		
77	31.4	24.6	37.0



#54
 cis-1,3-Dichloropropene
 Concen: 55.30 ug/l
 RT: 8.44 min Scan# 1195
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
75	206234		
77	31.6	26.2	39.2
39	40.4	36.4	54.6









#55
 1,1,2-Trichloroethane
 Concen: 50.22 ug/l
 RT: 9.22 min Scan# 1323
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Instrument : MSVOA_X
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
97	138776		
97	100		
83	85.6	66.4	99.6
85	55.8	43.3	64.9
99	63.0	49.0	73.4




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 53.52 ug/l
 RT: 9.18 min Scan# 1317
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
69	209245		
69	100		
41	62.0	51.0	76.4
39	28.5	26.1	39.1



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0









#57
 1,3-Dichloropropane
 Concen: 50.35 ug/l
 RT: 9.37 min Scan# 1348
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
76	226085		
78	32.3	25.5	38.3



#58
 2-Chloroethyl Vinyl ether
 Concen: 242.45 ug/l
 RT: 8.32 min Scan# 1175
 Delta R.T. 0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
63	482476		
106	28.6	23.9	35.9









#59
 2-Hexanone
 Concen: 246.88 ug/l
 RT: 9.50 min Scan# 1369
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 741175
 Ion Ratio Lower Upper
 43 100
 58 57.8 29.0 87.0



#60
 Dibromochloromethane
 Concen: 54.73 ug/l
 RT: 9.59 min Scan# 1383
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion: 129 Resp: 142389
 Ion Ratio Lower Upper
 129 100
 127 76.8 38.5 115.5









#61
 1,2-Dibromoethane
 Concen: 53.79 ug/l
 RT: 9.68 min Scan# 1398
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
107	141992		
109	95.0	76.0	114.0



#62
 4-Bromofluorobenzene
 Concen: 46.27 ug/l
 RT: 11.14 min Scan# 1638
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
95	153280		
174	94.6	0.0	185.2
176	92.2	0.0	178.6









#63
 Chlorobenzene-d5
 Concen: 50.00 ug/l
 RT: 10.12 min Scan# 1470
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
117	286433		
82	52.8	47.8	71.6
119	32.0	29.3	43.9



#64
 Tetrachloroethene
 Concen: 53.54 ug/l
 RT: 9.34 min Scan# 1342
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
164	191452		
166	130.7	102.8	154.2
129	90.2	69.4	104.0
131	86.9	67.9	101.9









#65
 Chlorobenzene
 Concen: 54.00 ug/l
 RT: 10.14 min Scan# 1474
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Instrument :
 MSVOA_X
 ClientSampled :

Tgt Ion: 112 Resp: 384172




Ion	Ratio	Lower	Upper
112	100		
114	32.0	27.8	41.8




#66
 1,1,1,2-Tetrachloroethane
 Concen: 56.28 ug/l
 RT: 10.23 min Scan# 1488
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion: 131 Resp: 138088

Ion	Ratio	Lower	Upper
131	100		
133	95.3	48.1	144.4
119	65.4	32.9	98.6









#67
 Ethyl Benzene
 Concen: 55.27 ug/l
 RT: 10.26 min Scan# 1493
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion: 91 Resp: 639463
 Ion Ratio Lower Upper
 91 100
 106 32.1 25.9 38.9



#68
 m/p-Xylenes
 Concen: 113.35 ug/l
 RT: 10.36 min Scan# 1510
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion: 106 Resp: 505114
 Ion Ratio Lower Upper
 106 100
 91 198.1 157.1 235.7









#69
 o-Xylene
 Concen: 57.47 ug/l
 RT: 10.70 min Scan# 1566
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
106	246200		
91	208.2	105.1	315.1



#70
 Styrene
 Concen: 58.58 ug/l
 RT: 10.71 min Scan# 1568
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
104	415766		
78	46.9	37.4	56.0
103	55.2	43.8	65.6









#71
 Bromoform
 Concen: 57.02 ug/l
 RT: 10.86 min Scan# 1592
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
173	114872		
175	49.0	24.1	72.3
254	0.2	0.2	0.2



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.00 ug/l
 RT: 12.08 min Scan# 1792
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
152	177116		
115	79.1	39.0	117.0
150	175.3	0.0	351.0









#73
 Isopropylbenzene
 Concen: 56.73 ug/l
 RT: 11.02 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
105	100		
120	26.9	13.5	40.4



#74
 N-ethyl acetate
 Concen: 53.96 ug/l
 RT: 10.90 min Scan# 1599
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
43	100		
70	45.2	37.4	56.0
55	26.2	21.8	32.8
61	25.6	20.6	30.8









#75
 1,1,2,2-Tetrachloroethane
 Concen: 50.22 ug/l
 RT: 11.27 min Scan# 1659
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
83	188955		
131	10.7	5.2	15.6
85	64.6	32.3	96.8



#76
 1,2,3-Trichloropropane
 Concen: 68.39 ug/l
 RT: 11.30 min Scan# 1664
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
75	243707		
77	32.0	20.2	60.6









#77
 Bromobenzene
 Concen: 55.97 ug/l
 RT: 11.26 min Scan# 1657
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
156	180445		
77	138.1	71.5	214.3
158	96.6	48.9	146.8



#78
 n-propylbenzene
 Concen: 56.05 ug/l
 RT: 11.36 min Scan# 1674
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
91	751693		
120	24.2	11.9	35.9









#79
 2-Chlorotoluene
 Concen: 54.36 ug/l
 RT: 11.42 min Scan# 1684
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
91	442149		
126	36.0	17.9	53.7



#80
 1,3,5-Trimethylbenzene
 Concen: 56.89 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
105	557906		
120	49.8	25.6	76.8









#81
 trans-1,4-Dichloro-2-butene
 Concen: 53.76 ug/l
 RT: 11.08 min Scan# 1628
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
75	100		
53	91.5	80.1	120.1
89	45.6	37.2	55.8



#82
 4-Chlorotoluene
 Concen: 54.64 ug/l
 RT: 11.51 min Scan# 1699
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
91	100		
126	31.8	15.5	46.5









#83
 tert-Butylbenzene
 Concen: 56.56 ug/l
 RT: 11.77 min Scan# 1742
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
119	527988		
91	100	27.0	81.0
134	100	11.7	35.1



#84
 1,2,4-Trimethylbenzene
 Concen: 57.12 ug/l
 RT: 11.81 min Scan# 1748
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
105	571917		
120	100	23.2	69.6









#85
 sec-Butylbenzene
 Concen: 57.06 ug/l
 RT: 11.95 min Scan# 1771
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
105	678769		
134	21.0	10.5	31.5



#86
 p-Isopropyltoluene
 Concen: 57.93 ug/l
 RT: 12.07 min Scan# 1790
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
119	608593		
134	26.8	13.4	40.2
91	21.5	10.9	32.7









#87
 1,3-Dichlorobenzene
 Concen: 54.72 ug/l
 RT: 12.03 min Scan# 1784
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
146	100		
111	37.1	18.7	56.1
148	64.3	32.1	96.5



#88
 1,4-Dichlorobenzene
 Concen: 53.37 ug/l
 RT: 12.10 min Scan# 1795
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
146	100		
111	36.0	18.1	54.1
148	64.1	32.0	96.0









#89
 n-Butylbenzene
 Concen: 54.51 ug/l
 RT: 12.39 min Scan# 1843
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
91	521800		
92	53.1	27.3	81.9
134	27.6	13.6	40.7



#90
 Hexachloroethane
 Concen: 56.55 ug/l
 RT: 12.60 min Scan# 1877
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
117	95099		
117	100		
201	100.7	46.9	140.6









#91
 1,2-Dichlorobenzene
 Concen: 55.16 ug/l
 RT: 12.40 min Scan# 1844
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument : MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
146	100		
111	38.2	19.4	58.1
148	63.9	32.8	98.3



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 56.44 ug/l
 RT: 13.00 min Scan# 1943
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
75	100		
155	100.7	48.0	144.2
157	128.4	61.4	184.1









#93
 1,2,4-Trichlorobenzene
 Concen: 58.13 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
180	100		
182	95.3	48.4	145.0
145	27.4	14.3	42.9



#94
 Hexachlorobutadiene
 Concen: 54.88 ug/l
 RT: 13.79 min Scan# 2072
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
225	100		
223	61.7	30.1	90.5
227	63.7	30.8	92.4









#95
 Naphthalene
 Concen: 60.49 ug/l
 RT: 13.83 min Scan# 2080
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Instrument : MSVOA_X
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
128	731376		
127	12.7	10.2	15.2
129	10.8	8.6	12.8



#96
 1,2,3-Trichlorobenzene
 Concen: 57.70 ug/l
 RT: 14.02 min Scan# 2111
 Delta R.T. -0.00 min
 Lab File: VX004719.D
 Acq: 19 Sep 2018 21:13

Tgt Ion	Resp	Lower	Upper
180	255542		
182	95.4	48.5	145.6
145	29.2	14.9	44.9

