

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004707.D
 Acq On : 19 Sep 2018 15:52
 Operator : JC/MD
 Sample : J4939-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Sep 20 13:06:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	215580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	340383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166970	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	143965	51.79	ug/l	0.00
Spiked Amount			Recovery	=	103.58%	
35) Dibromofluoromethane	5.50	113	123698	39.71	ug/l	0.00
Spiked Amount			Recovery	=	79.42%	
50) Toluene-d8	8.71	98	455521	43.61	ug/l	0.00
Spiked Amount			Recovery	=	87.22%	
62) 4-Bromofluorobenzene	11.14	95	158526	42.52	ug/l	0.00
Spiked Amount			Recovery	=	85.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	29752	10.43	ug/l	99
12) 1,1-Dichloroethene	2.37	96	194413	86.31	ug/l	94
16) Acetone	2.45	43	3710	2.73	ug/l	89
24) 1,1-Dichloroethane	3.70	63	143945	30.57	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	41368	14.89	ug/l	89
42) 1,2-Dichloroethane	6.20	62	9081	2.23	ug/l	99
44) Trichloroethene	7.22	130	3174	1.00	ug/l	86
49) 1,4-Dioxane	7.75	88	43363	593.04	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091918\
Data File : VX004707.D
Acq On : 19 Sep 2018 15:52
Operator : JC/MD
Sample : J4939-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Sep 20 13:06:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

