

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004713.D
 Acq On : 19 Sep 2018 18:33
 Operator : JC/MD
 Sample : J4975-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-220

Quant Time: Sep 20 04:45:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	218900	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	335375	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	318085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	137261	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	5.50	113	125073	40.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.50%	
50) Toluene-d8	8.71	98	470145	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
62) 4-Bromofluorobenzene	11.14	95	166827	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
Target Compounds						
16) Acetone	2.44	43	44741	32.48	ug/l	98
18) Methyl Acetate	2.78	43	5408	1.40	ug/l	94
20) Methylene Chloride	2.85	84	17053	6.03	ug/l	98
25) 2-Butanone	4.70	43	18216	8.31	ug/l	100
43) Isopropyl Acetate	6.47	43	133639	21.73	ug/l	99
95) Naphthalene	13.83	128	36699	3.08	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091918\
Data File : VX004713.D
Acq On : 19 Sep 2018 18:33
Operator : JC/MD
Sample : J4975-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-220

Quant Time: Sep 20 04:45:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

