

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004718.D
 Acq On : 19 Sep 2018 20:46
 Operator : JC/MD
 Sample : J4916-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FC-MW1011-20180911

Quant Time: Sep 20 05:01:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	216125	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	326064	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	311420	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	177020	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	132743	47.63	ug/l	0.00
Spiked Amount			Recovery	=	95.26%	
35) Dibromofluoromethane	5.50	113	125450	42.04	ug/l	0.00
Spiked Amount			Recovery	=	84.08%	
50) Toluene-d8	8.71	98	455868	45.56	ug/l	0.00
Spiked Amount			Recovery	=	91.12%	
62) 4-Bromofluorobenzene	11.14	95	163469	45.77	ug/l	0.00
Spiked Amount			Recovery	=	91.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	3684	2.71	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091918\
Data File : VX004718.D
Acq On : 19 Sep 2018 20:46
Operator : JC/MD
Sample : J4916-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FC-MW10I1-20180911

Quant Time: Sep 20 05:01:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

